

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2018]玛社量标 证字 第06号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 玛纳斯县质量与计量检测所

Possessor of the measurement standard

计量标准名称 精密压力表标准装置

代码 12415100

Name of measurement standard

Code

测量范围 (-0.1-60) MPa

Measuring range

不确定度或准确度等级

或最大允许误差 0.4级

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章) 玛纳斯县质量技术监督局

The issuing authority

发证日期 2018年11月01日

Date issued

有效期至 2022年10月31日

Date of expiry

[2018]玛社量标 证字 第06号

| 可开展检定及校准 项目名称 Verification and calibration items | 测量范围 Measuring range | 不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error |
|--|-------------------------|---|
| 压力表 氧气表 | (-0.1-60) MPa | 1.6 级及以下 |
| 传递地区或范围 Region of service | 玛纳斯县行政区域内 | |
| 计量标准考核证书号 Certificate No. (for examination of measurement standard) | [2018]昌州量标 证字第 14 号 | |
| 计量授权证书号 Certificate No. (for authorization) | (新)法计(2014)县级 027 号 | |