

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2016]玛社量标 证字 第 201601 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 玛纳斯县质量与计量检测所

Possessor of the measurement standard

计量标准名称 数字三用表检定装置

代码 15314507

Name of measurement standard

Code

测量范围 DCV, ACV: (0-1) KV;

DCI, ACI(0-10)A;

Ω : 10 Ω 100M Ω

Measuring range

不确定度或准确度等级

或最大允许误差 0.1 级

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章) 玛纳斯县质量技术监督局

The issuing authority

发证日期 2016 年 05 月 15 日

Date issued

有效期至 2020 年 05 月 14 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
交直流电压表 交直流电流表	DCV, ACV (0-1)KV DCI, ACI (0-10)A	0.5 级及以下
传递地区或范围 Region of service	玛纳斯县行政区域内	
计量标准考核证书号 Certificate No. (for examination of measurement standard)	[2016]昌州量标 证字第 201601 号	
计量授权证书号 Certificate No. (for authorization)	(新)法计(2014)县级 027 号	